

Dr Baba Saheb Ambedkar Hospital  
(Govt of NCT of Delhi)  
Sector-6, Rohini, Delhi – 110085

Hospital Open Tender – List of Items (O&G)

Tender ID – 2017\_BSAH\_130983\_6

S No	Name of the items	Specification If Any	Unit(s) required*
1	OFFICE HYSTEROSCOPE	<ul style="list-style-type: none"> <li>➤ Forward-oblique telescope 30 degree, autoclavable with fiberoptic light connections, dia 2.9mm, length 30 cm.</li> <li>➤ Light cable 200-230cm. Compatible with the Hysteroscope.</li> <li>➤ Compatible operative sheath with continuous flow channel.</li> <li>➤ One scissor</li> <li>➤ One grasper and</li> <li>➤ One biopsy forceps 5fr, length 34-35cm.</li> <li>➤ Should be CE/FDA approved / ISO certified Equipment should be maintained under CMC after the warranty period</li> </ul>	1 ( ONE) Nos.
2	COLPOSCOPE	<ul style="list-style-type: none"> <li>➤ For visualization of cervix and vagina, video colposcope, automatic focus up to 30 cm distance from the patient, Green filter, Zoom 1-40x, Super bright shadow less light, Real time display, Recording and image saving facility, Computer monitor, Foot operated and computer operated.</li> <li>➤ CE/ISO certified</li> </ul>	1 ( ONE) Nos.
3	LABOUR TABLE SST WITH MATTRESS WITH ALL STANDARD ACCESSORIES	<ul style="list-style-type: none"> <li>➤ Labour table with thick stainless steel top (noncoated) with removable mattress with a all standard accessories.</li> <li>➤ Overall approximate Dimension 70-75" L x 27-35" W x 30-32" H</li> <li>➤ Three section top</li> <li>➤ Backrest adjustable on ratchet</li> <li>➤ Middle section with "U" cut</li> <li>➤ Leg section can be telescoped under the main frame completely for lithotomy positions</li> <li>➤ Should have facility for Trendelenberg's and reverse trendelenberg's position</li> </ul>	3 (THREE) Nos.

Dr Baba Saheb Ambedkar Hospital  
(Govt of NCT of Delhi)  
Sector-6, Rohini, Delhi – 110085

Hospital Open Tender – List of Items (O&G)

Tender ID – 2017\_BSAH\_130983\_6

		<ul style="list-style-type: none"> <li>➤ Should have drop down Stainless steel railing on three sides (head and side railing)</li> <li>➤ Should have hand grips, a pair of rubber padded knee crutches with Stainless steel lithotomy poles and Stainless steel bowl.</li> <li>➤ Mattress should be equal / more than 25 mm thickness and cover fabric should be hygienic, durable and washable.</li> <li>➤ Rate for extra mattress should be quoted separately.</li> </ul>	
<b>4</b>	LABOUR BED CUM DELIVERY TABLE	<ul style="list-style-type: none"> <li>➤ Electrically operated 3 sectional bed to be converted into multifunctional bed for obstetrical patients during stages of labour approximate dimensions.</li> <li>➤ full length 1800-2000mm; width 660-700mm; height 600-800mm</li> <li>➤ Weight bearing capacity approx. 130 kgs reverse.</li> <li>➤ Trendelenberg / reverse Trendelenberg positions, height adjustment, back section adjustment through electronic system</li> <li>➤ Bed equipped with side rails knee crutches 2, foot support 2, stainless steel basin, infusion rod.</li> <li>➤ Mattress cover fabric should be hygienic, durable and washable and thick for patient comfort.</li> <li>➤ Equipment should be maintained under CMC after the warranty period</li> <li>➤ CE/ISO certified</li> </ul>	3 (THREE) Nos.

\*The quantity is approximate and may vary at the time of placing orders.